



PROPERTY EQUIPMENT CARD

A

ANALYSIS

FNC-A-023

MEASURING NAME	TURBIDITY METER
----------------	-----------------

SPECIFICATION

1	MODEL	HI93414	HANNA / IT
2	MEASUREMENT RANGE	0~1000 NTU	
3	VALUE ACCURACY	±2% +0.02 NTU	
4	REPETITION	±1% 0.02 NTU	
5	POWER	4 x 1.5V AA 12Vdc	

EQUIPMENT IMAGE

